

**Ref: FAQ0128****Version: 1.0****Title – Hydrins material content details****Made By: AB 15/09/15****(Issue 2)**

Hydrins II - material content details

All the following parts are Stainless Steel grade SS316:

- Sensor tip,
- Sensor electrodes (x2),
- Sensor body,
- Stem,
- Gland body.

The following parts made of Brass grade CZ121:

- Lower bush,
- Support bush

The following part is made of Nickel plated brass :

- Pressure fitting.

All synthetic parts are WRAS approved as indicated in the documentation and Hydreka are currently working seeking accreditation to ACS (Attestation de Conformité Sanitaire) for France which is a very demanding criteria regulation for the Hydrins insertion probe.

Document History:

Edition	Date of Issue	Modification	Notes
1st	03/09/12	Release	
2nd	15/09/15	Format update	