

**Ref: FAQ0193****Version: 1.0****Title – Mutlilog & LX low battery alarms****Made By: AB 17/09/15****(Issue 2)**

Mutlilog and LX Low battery alarms

Mutlilog data loggers have 2 alarms, one triggered by the logger battery and one triggered by the amount of time the modem has been on and the number of GSM batteries fitted. The modem 'on' time is not registered when external power is supplied so a single alarm is only monitoring the internal logger batteries.

Mutlilog LX has a single battery alarm based on battery voltage. This is sent every 2 days via SMS when the low threshold is reached. It is sent to the first call out number.

Low battery alarms should be active for around 14 days prior to the battery depleting.

Document History:

Edition	Date of Issue	Modification	Notes
1st		Release	
2nd	17/09/15	Format update	