



Ref: FAQ0121

Version: 1.0

Title-Radcom loggers Ext batt box part #

Made By: AB 15/09/15

(Issue 2)

## Radcom Loggers - Specs and part numbers of the external battery boxes

External Battery Requirement for 5 years Part Number	Description of Battery Box
EXTBATTBOX60/4	4pin connector for LX Approx 5yr Battery Life @1 Hour Dial In
EXTBATTBOX60/10	10pin connector for Multilog Approx 5yr Battery Life @1 Hour Dial In
EXTBATTBOX30/4	4pin connector for LX Approx 5yr Battery Life @30 Minute Dial In
EXTBATTBOX30/10	10pin connector for Multilog Approx 5yr Battery Life @30 Minute Dial In
EXTBATTBOX15/4	4pin connector for LX Approx 5yr Battery Life @15 Minute Dial In
EXTBATTBOX15/10	10pin connector for Multilog Approx 5yr Battery Life @15 Minute Dial In

### Document History:

Edition	Date of Issue	Modification	Notes
1st	24/05/12	Release	
2nd	15/09/15	Format update	